P645 SPECIALITE

DWG3
Unit ID:



NOTE: HANDING IS DETERMINED AS PATIENT LIES FACE UP IN BED.



TOTAL QUANTITY: _____

TYPICAL ILLUSTRATION ONLY
DO NOT Use For Construction

INFORMATION:

OVERALL LENGTH:	☐ 37 9/16"	☐ 49 9/16"	
POWER SOURCE:	☐ NORMAL POWER	☐ CRITICAL BRANCH	
VOLTAGE:	☐ 120V	☐ 220/240V	☐ 277V

HPL TRIM COLOR: _____

(Optional):

FUSED BALLAST:	☐ YES	☐ NO	
CHART LIGHT/DUPLEX:	☐ RIGHT SIDE	☐ LEFT SIDE	
LOW VOLTAGE CONTROLLER:	☐ 120V	☐ 220/240V	☐ 277V
REMOTE LIGHT SWITCHING REQUIRED:	☐ INDIRECT	☐ READ	

NOTE: REMOTE SWITCHING OF READ LIGHT NOT AVAILABLE IF USED IN CONJUNCTION WITH PULL CORD SWITCH

PULL CORD SWITCH (One per light):	☐ RIGHT HAND	☐ CENTER	☐ LEFT HAND

NOTE: PULL CORD SWITCH NOT AVAILABLE IN CONJUNCTION W/ 277V OR LOW VOLTAGE CONTROLLER
ALL REMOTE SWITCHES PROVIDE BY OTHERS

CUSTOMER #: _____ *PURCHASE ORDER #:* _____

HOSPITAL or COMPANY: _____

ADDRESS: _____

CITY: _____ *STATE:* _____ *ZIP:* _____

PHONE #: _____ *FAX #:* _____

TOTAL QUANTITY: _____ *REQUESTED DELIVERY DATE:* _____

HILL-ROM ORDER #: _____ *HILL-ROM FAX #:* 812-934-1697

CUSTOMER APPROVAL SIGNATURE REQUIRED BELOW.

Please complete this room handing chart.
List the room#'s and check the desired unit handing.

Room #	As Shown	Opposite (Mirror Image)

IMPORTANT:
- PLEASE NOTE THAT ANY DELAYS IN RETURNING THE APPROVED SUBMITTALS, OR RETURNING THEM WITH INCOMPLETE OR INACCURATE INFORMATION, *WILL CAUSE DELAYS IN DELIVERY* OR POSSIBLY UNITS BUILT WITH WRONG ELECTRICAL AND MECHANICAL SERVICES.
- PLEASE VERIFY THAT THE ABOVE INFORMATION (WHERE ALREADY PROVIDED) IS CORRECT, AND SUPPLY THE REMAINING INFORMATION NECESSARY TO PROCESS THE ORDER. ALSO, PLEASE BE SURE TO PROVIDE THE APPROPRIATE APPROVAL SIGNATURE(S) AS AUTHORIZATION TO PROCEED WITH MANUFACTURING.
- NORMAL DELIVERY IS 4 TO 6 WEEKS AFTER RECEIPT OF APPROVED, SIGNED SUBMITTAL DRAWINGS.

_____ _____ _____
APPROVAL SIGNATURE DATE PHONE

REQUESTED DELIVERY DATE: _____

This Information Sheet Applies to the Following Drawing(s):

NOTICE:
The drawings and all information thereon is property of:
HILL-ROM COMPANY, Inc.
Batesville, Indiana
This drawing and information thereon, is confidential; and must not be made public, copied or used to disadvantage of said Hill-Rom Company, Inc. It is subject to return on demand.

Architectural Products
1069 State Route 46 East
Batesville, Indiana 47006
Ph: (812)934 7777
http:\\www.hill-rom.com

Hill-Rom®
A HILLENBRAND INDUSTRY

Project Manager:
CPT: PE:
Drawn By: StimS Checked By:
Room Description:

Customer:
PRICE BOOK
TYPICAL SPECIALITE 645-00
DRAWING #3 / 2005

Customer ID:
Quantity: 1

Date: 01/01/04 Scale:
Request For Information
File Name: PBDRAWING3